

Surrogate Summary

SDG No.: **P4600**

Client: **Tetra Tech NUS, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P4600-01	BP-VPB-190-TB-20241025	1,2-Dichloroethane-d4	50	43.6	87	81	118
		Dibromofluoromethane	50	45.3	91	80	119
		Toluene-d8	50	50.0	100	89	112
		4-Bromofluorobenzene	50	46.5	93	85	114
P4600-02	BP-VPB-190-DUP-20241025	1,2-Dichloroethane-d4	50	44.0	88	81	118
		Dibromofluoromethane	50	45.8	92	80	119
		Toluene-d8	50	49.9	100	89	112
		4-Bromofluorobenzene	50	45.4	91	85	114
P4600-03	BP-VPB-190-GW-438-440	1,2-Dichloroethane-d4	50	43.3	87	81	118
		Dibromofluoromethane	50	45.2	90	80	119
		Toluene-d8	50	49.9	100	89	112
		4-Bromofluorobenzene	50	48.2	96	85	114
P4600-04	BP-VPB-190-GW-458-460	1,2-Dichloroethane-d4	50	47.1	94	81	118
		Dibromofluoromethane	50	46.9	94	80	119
		Toluene-d8	50	51.4	103	89	112
		4-Bromofluorobenzene	50	51.1	102	85	114
P4600-05	BP-VPB-190-GW-478-480	1,2-Dichloroethane-d4	50	42.0	84	81	118
		Dibromofluoromethane	50	46.0	92	80	119
		Toluene-d8	50	50.0	100	89	112
		4-Bromofluorobenzene	50	47.0	94	85	114
P4600-06	BP-VPB-190-GW-498-500	1,2-Dichloroethane-d4	50	43.8	88	81	118
		Dibromofluoromethane	50	45.9	92	80	119
		Toluene-d8	50	50.0	100	89	112
		4-Bromofluorobenzene	50	51.2	102	85	114
P4600-07	BP-VPB-190-GW-518-520	1,2-Dichloroethane-d4	50	42.8	85	81	118
		Dibromofluoromethane	50	45.2	90	80	119
		Toluene-d8	50	50.5	101	89	112
		4-Bromofluorobenzene	50	48.3	97	85	114
VX1029WBL01	VX1029WBL01	1,2-Dichloroethane-d4	50	43.4	87	81	118
		Dibromofluoromethane	50	45.5	91	80	119
		Toluene-d8	50	50.5	101	89	112
		4-Bromofluorobenzene	50	50.6	101	85	114
VX1029WBS01	VX1029WBS01	1,2-Dichloroethane-d4	50	42.1	84	81	118
		Dibromofluoromethane	50	46.6	93	80	119
		Toluene-d8	50	48.3	97	89	112
		4-Bromofluorobenzene	50	47.0	94	85	114
VX1029WBSD0	VX1029WBSD01	1,2-Dichloroethane-d4	50	42.9	86	81	118
		Dibromofluoromethane	50	47.2	94	80	119
		Toluene-d8	50	50.0	100	89	112
		4-Bromofluorobenzene	50	50.0	100	85	114
VX1030WBL01	VX1030WBL01	1,2-Dichloroethane-d4	50	44.0	88	81	118
		Dibromofluoromethane	50	45.0	90	80	119
		Toluene-d8	50	50.0	100	89	112
		4-Bromofluorobenzene	50	47.0	94	85	114
VX1030WBS01	VX1030WBS01	1,2-Dichloroethane-d4	50	46.0	92	81	118
		Dibromofluoromethane	50	48.0	96	80	119
		Toluene-d8	50	51.0	102	89	112
		4-Bromofluorobenzene	50	52.0	104	85	114