

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: P4600 SAS No.: P4600 SDG NO.: P4600
Lab File ID: VX043583.D Date Analyzed: 10/29/2024
Instrument ID: MSVOA_X Time Analyzed: 09:29
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	136145	12.018				
UPPER LIMIT	272290	12.518				
LOWER LIMIT	68072.5	11.518				
EPA SAMPLE NO.						
BP-VPB-190-TB-20241025	104072	12.02				
BP-VPB-190-DUP-20241025	109255	12.02				
BP-VPB-190-GW-438-440	105995	12.02				
BP-VPB-190-GW-458-460	83818	12.02				
BP-VPB-190-GW-498-500	99219	12.02				
BP-VPB-190-GW-518-520	98034	12.02				
VX1029WBL01	128354	12.02				
VX1029WBS01	142936	12.02				
VX1029WBSD01	131226	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.