

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: P4600 SAS No.: P4600 SDG NO.: P4600
Lab File ID: VX043582.D BFB Injection Date: 10/29/2024
Instrument ID: MSVOA_X BFB Injection Time: 08:30
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	16.5
75	30.0 - 60.0% of mass 95	50.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6
173	Less than 2.0% of mass 174	0.5 (0.7) 1
174	50.0 - 100.0% of mass 95	73.3
175	5.0 - 9.0% of mass 174	5.5 (7.5) 1
176	95.0 - 101.0% of mass 174	70.5 (96.2) 1
177	5.0 - 9.0% of mass 176	4.7 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX043583.D	10/29/2024	09:29
VX1029WBL01	VX1029WBL01	VX043585.D	10/29/2024	10:31
VX1029WBS01	VX1029WBS01	VX043586.D	10/29/2024	10:57
VX1029WBSD01	VX1029WBSD01	VX043587.D	10/29/2024	11:20
BP-VPB-190-TB-20241025	P4600-01	VX043599.D	10/29/2024	16:01
BP-VPB-190-DUP-20241025	P4600-02	VX043600.D	10/29/2024	16:24
BP-VPB-190-GW-438-440	P4600-03	VX043601.D	10/29/2024	16:48
BP-VPB-190-GW-458-460	P4600-04	VX043602.D	10/29/2024	17:11
BP-VPB-190-GW-498-500	P4600-06	VX043604.D	10/29/2024	17:57
BP-VPB-190-GW-518-520	P4600-07	VX043605.D	10/29/2024	18:20
VSTDCCC050EC	VSTDCCC050	VX043608.D	10/29/2024	19:29