

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: TETR16

Lab Code: CHEM Case No.: P4602 SAS No.: P4602 SDG NO.: P4602

Instrument ID: ECD_S Calibration Date(s): 11/06/2024 11/06/2024
Calibration Times: 09:48 11:24

GC Column: RTX-CLP ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS028253.D</u>	CF 500 =	<u>PS028254.D</u>		
CF 750 =		<u>PS028255.D</u>	CF 1000 =	<u>PS028256.D</u>	CF 1500 =	<u>PS028257.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-T	18605000000	17699100000	17397600000	16764700000	16211200000	17335500000	5
2,4,5-TP(Silvex)	18381400000	17393400000	17071400000	16437600000	15844400000	17025600000	6
2,4-D	3499520000	3198390000	3124710000	3013570000	2944350000	3156110000	7
2,4-DB	2953740000	2827030000	2824430000	2764350000	2763120000	2826530000	3
2,4-DCAA	2911600000	2530960000	2546610000	2432090000	2385800000	2561410000	8
DICAMBA	11781800000	11346100000	11291200000	10963200000	10760300000	11228500000	3
DICHLORPROP	3148380000	2853420000	2784160000	2683190000	2627250000	2819280000	7
Dinoseb	15121200000	14574900000	14554700000	14126500000	13917700000	14459000000	3

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: TETR16

Lab Code: CHEM Case No.: P4602 SAS No.: P4602 SDG NO.: P4602

Instrument ID: ECD_S Calibration Date(s): 11/06/2024 11/06/2024
Calibration Times: 09:48 11:24

GC Column: RTX-CLP2 ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS028253.D</u>	CF 500 =	<u>PS028254.D</u>		
CF 750 =		<u>PS028255.D</u>	CF 1000 =	<u>PS028256.D</u>	CF 1500 =	<u>PS028257.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-T	12788400000	12436700000	12406800000	12012600000	11689500000	12266800000	3
2,4,5-TP(Silvex)	12950900000	12612000000	12578800000	12192000000	11879700000	12442700000	3
2,4-D	2348170000	2233230000	2226180000	2175820000	2167290000	2230140000	3
2,4-DB	1594690000	1565870000	1590090000	1575360000	1598480000	1584900000	1
2,4-DCAA	1855960000	1741220000	1731630000	1690270000	1689820000	1741780000	4
DICAMBA	7963390000	8019410000	8155900000	8026840000	7997280000	8032560000	1
DICHLORPROP	2102670000	2006780000	2002310000	1963630000	1970820000	2009240000	3
Dinoseb	8938980000	8771400000	8800740000	8588540000	8508600000	8721650000	2