

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: P4602 SAS No.: P4602 SDG NO.: P4602

Continuing Calib Date: 11/06/2024 Initial Calibration Date(s): 11/05/2024 11/05/2024

Continuing Calib Time: 01:24 Initial Calibration Time(s): 17:47 19:23

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.45	7.46	7.36	7.56	0.01
2,4-DCAA	7.27	7.27	7.17	7.37	0.01
DICHLORPROP	8.17	8.17	8.07	8.27	0.00
2,4-D	8.40	8.40	8.30	8.50	0.00
2,4,5-TP(Silvex)	9.28	9.28	9.18	9.38	0.00
2,4,5-T	9.57	9.57	9.47	9.67	0.00
2,4-DB	10.15	10.15	10.05	10.25	0.00
Dinoseb	11.37	11.37	11.27	11.47	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: P4602 SAS No.: P4602 SDG NO.: P4602

Continuing Calib Date: 11/06/2024 Initial Calibration Date(s): 11/05/2024 11/05/2024

Continuing Calib Time: 01:24 Initial Calibration Time(s): 17:47 19:23

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.97	7.97	7.87	8.07	0.00
2,4-DCAA	7.77	7.77	7.67	7.87	0.00
DICHLORPROP	8.69	8.69	8.59	8.79	0.00
2,4-D	9.02	9.02	8.92	9.12	0.00
2,4,5-TP(Silvex)	9.93	9.93	9.83	10.03	0.00
2,4,5-T	10.35	10.35	10.25	10.45	0.00
2,4-DB	10.92	10.92	10.82	11.02	0.01
Dinoseb	11.30	11.30	11.20	11.40	0.00