

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: P4602 SAS No.: P4602 SDG NO.: P4602

Continuing Calib Date: 11/07/2024 Initial Calibration Date(s): 11/06/2024 11/06/2024

Continuing Calib Time: 04:40 Initial Calibration Time(s): 09:48 11:24

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.45	7.45	7.35	7.55	0.00
2,4-DCAA	7.26	7.26	7.16	7.36	0.00
DICHLORPROP	8.16	8.16	8.06	8.26	0.00
2,4-D	8.39	8.39	8.29	8.49	0.00
2,4,5-TP(Silvex)	9.27	9.28	9.18	9.38	0.01
2,4,5-T	9.56	9.57	9.47	9.67	0.01
2,4-DB	10.14	10.15	10.05	10.25	0.01
Dinoseb	11.36	11.36	11.26	11.46	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: P4602 SAS No.: P4602 SDG NO.: P4602

Continuing Calib Date: 11/07/2024 Initial Calibration Date(s): 11/06/2024 11/06/2024

Continuing Calib Time: 04:40 Initial Calibration Time(s): 09:48 11:24

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.96	7.97	7.87	8.07	0.01
2,4-DCAA	7.76	7.77	7.67	7.87	0.01
DICHLORPROP	8.68	8.68	8.58	8.78	0.00
2,4-D	9.01	9.02	8.92	9.12	0.01
2,4,5-TP(Silvex)	9.92	9.92	9.82	10.02	0.00
2,4,5-T	10.34	10.34	10.24	10.44	0.00
2,4-DB	10.91	10.91	10.81	11.01	0.00
Dinoseb	11.29	11.29	11.19	11.39	0.00