

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: P4602 SAS No.: P4602 SDG NO.: P4602

Continuing Calib Date: 11/09/2024 Initial Calibration Date(s): 11/08/2024 11/08/2024

Continuing Calib Time: 00:34 Initial Calibration Time(s): 15:13 16:58

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.44	7.44	7.34	7.54	0.00
2,4-DCAA	7.25	7.25	7.15	7.35	0.00
DICHLORPROP	8.15	8.15	8.05	8.25	0.00
2,4-D	8.38	8.38	8.28	8.48	0.00
2,4,5-TP(Silvex)	9.26	9.26	9.16	9.36	0.00
2,4,5-T	9.55	9.55	9.45	9.65	0.00
2,4-DB	10.13	10.13	10.03	10.23	0.00
Dinoseb	11.34	11.35	11.25	11.45	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: P4602 SAS No.: P4602 SDG NO.: P4602

Continuing Calib Date: 11/09/2024 Initial Calibration Date(s): 11/08/2024 11/08/2024

Continuing Calib Time: 00:34 Initial Calibration Time(s): 15:13 16:58

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.95	7.96	7.86	8.06	0.01
2,4-DCAA	7.75	7.76	7.66	7.86	0.01
DICHLORPROP	8.67	8.67	8.57	8.77	0.00
2,4-D	9.00	9.01	8.91	9.11	0.01
2,4,5-TP(Silvex)	9.91	9.91	9.81	10.01	0.00
2,4,5-T	10.33	10.33	10.23	10.43	0.00
2,4-DB	10.90	10.90	10.80	11.00	0.00
Dinoseb	11.28	11.28	11.18	11.38	0.00