

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: P4602 SAS No.: P4602 SDG NO.: P4602

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/05/2024 11/05/2024

Client Sample No.: CCAL02 Date Analyzed: 11/06/2024

Lab Sample No.: HSTDCCC750 Data File : PS028237.D Time Analyzed: 01:24

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.572	9.473	9.673	925.990	712.500	30.0
2,4,5-TP(Silvex)	9.279	9.181	9.381	929.580	712.500	30.5
2,4-D	8.397	8.298	8.498	862.880	705.000	22.4
2,4-DB	10.150	10.051	10.251	933.840	712.500	31.1
2,4-DCAA	7.265	7.166	7.366	976.190	750.000	30.2
DICAMBA	7.454	7.355	7.555	836.620	705.000	18.7
DICHLORPROP	8.166	8.066	8.266	857.570	705.000	21.6
Dinoseb	11.367	11.268	11.468	1004.200	705.000	42.4

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: P4602 SAS No.: P4602 SDG NO.: P4602

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/05/2024 11/05/2024

Client Sample No.: CCAL02 Date Analyzed: 11/06/2024

Lab Sample No.: HSTDCCC750 Data File : PS028237.D Time Analyzed: 01:24

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.347	10.248	10.448	710.290	712.500	-0.3
2,4,5-TP(Silvex)	9.925	9.826	10.026	718.680	712.500	0.9
2,4-D	9.018	8.919	9.119	726.720	705.000	3.1
2,4-DB	10.915	10.815	11.015	689.350	712.500	-3.2
2,4-DCAA	7.767	7.667	7.867	788.490	750.000	5.1
DICAMBA	7.969	7.869	8.069	753.030	705.000	6.8
DICHLORPROP	8.686	8.587	8.787	733.990	705.000	4.1
Dinoseb	11.295	11.196	11.396	701.600	705.000	-0.5