

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: P4602 SAS No.: P4602 SDG NO.: P4602

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/06/2024 11/06/2024

Client Sample No.: CCAL03 Date Analyzed: 11/06/2024

Lab Sample No.: HSTDCCC750 Data File : PS028272.D Time Analyzed: 18:15

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.569	9.469	9.669	753.360	712.500	5.7
2,4,5-TP(Silvex)	9.276	9.177	9.377	748.270	712.500	5.0
2,4-D	8.393	8.294	8.494	733.030	705.000	4.0
2,4-DB	10.146	10.047	10.247	747.020	712.500	4.8
2,4-DCAA	7.261	7.162	7.362	777.510	750.000	3.7
DICAMBA	7.450	7.351	7.551	735.930	705.000	4.4
DICHLORPROP	8.162	8.063	8.263	731.120	705.000	3.7
Dinoseb	11.363	11.264	11.464	750.000	705.000	6.4

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: P4602 SAS No.: P4602 SDG NO.: P4602

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/06/2024 11/06/2024

Client Sample No.: CCAL03 Date Analyzed: 11/06/2024

Lab Sample No.: HSTDCCC750 Data File : PS028272.D Time Analyzed: 18:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.343	10.244	10.444	729.980	712.500	2.5
2,4,5-TP(Silvex)	9.922	9.823	10.023	740.260	712.500	3.9
2,4-D	9.014	8.915	9.115	721.950	705.000	2.4
2,4-DB	10.911	10.812	11.012	719.940	712.500	1.0
2,4-DCAA	7.764	7.665	7.865	776.030	750.000	3.5
DICAMBA	7.966	7.867	8.067	748.120	705.000	6.1
DICHLORPROP	8.683	8.584	8.784	731.010	705.000	3.7
Dinoseb	11.291	11.191	11.391	715.080	705.000	1.4