

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: P4602 SAS No.: P4602 SDG NO.: P4602

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/06/2024 11/06/2024

Client Sample No.: CCAL04 Date Analyzed: 11/06/2024

Lab Sample No.: HSTDCCC750 Data File : PS028284.D Time Analyzed: 23:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.567	9.469	9.669	767.920	712.500	7.8
2,4,5-TP(Silvex)	9.274	9.177	9.377	764.460	712.500	7.3
2,4-D	8.391	8.294	8.494	746.130	705.000	5.8
2,4-DB	10.144	10.047	10.247	761.720	712.500	6.9
2,4-DCAA	7.260	7.162	7.362	788.410	750.000	5.1
DICAMBA	7.449	7.351	7.551	747.530	705.000	6.0
DICHLORPROP	8.160	8.063	8.263	745.000	705.000	5.7
Dinoseb	11.361	11.264	11.464	760.470	705.000	7.9

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: P4602 SAS No.: P4602 SDG NO.: P4602

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/06/2024 11/06/2024

Client Sample No.: CCAL04 Date Analyzed: 11/06/2024

Lab Sample No.: HSTDCCC750 Data File : PS028284.D Time Analyzed: 23:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.342	10.244	10.444	716.510	712.500	0.6
2,4,5-TP(Silvex)	9.921	9.823	10.023	729.370	712.500	2.4
2,4-D	9.013	8.915	9.115	714.340	705.000	1.3
2,4-DB	10.909	10.812	11.012	703.300	712.500	-1.3
2,4-DCAA	7.763	7.665	7.865	773.710	750.000	3.2
DICAMBA	7.964	7.867	8.067	747.660	705.000	6.1
DICHLORPROP	8.682	8.584	8.784	725.580	705.000	2.9
Dinoseb	11.290	11.191	11.391	697.660	705.000	-1.0