

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: P4602 SAS No.: P4602 SDG NO.: P4602

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/06/2024 11/06/2024

Client Sample No.: CCAL05 Date Analyzed: 11/07/2024

Lab Sample No.: HSTDCCC750 Data File : PS028297.D Time Analyzed: 04:40

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.563	9.469	9.669	769.290	712.500	8.0
2,4,5-TP(Silvex)	9.270	9.177	9.377	773.230	712.500	8.5
2,4-D	8.388	8.294	8.494	749.690	705.000	6.3
2,4-DB	10.141	10.047	10.247	767.890	712.500	7.8
2,4-DCAA	7.258	7.162	7.362	772.740	750.000	3.0
DICAMBA	7.447	7.351	7.551	749.460	705.000	6.3
DICHLORPROP	8.157	8.063	8.263	750.990	705.000	6.5
Dinoseb	11.358	11.264	11.464	757.790	705.000	7.5

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: P4602 SAS No.: P4602 SDG NO.: P4602

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/06/2024 11/06/2024

Client Sample No.: CCAL05 Date Analyzed: 11/07/2024

Lab Sample No.: HSTDCCC750 Data File : PS028297.D Time Analyzed: 04:40

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.340	10.244	10.444	706.210	712.500	-0.9
2,4,5-TP(Silvex)	9.918	9.823	10.023	718.100	712.500	0.8
2,4-D	9.011	8.915	9.115	700.360	705.000	-0.7
2,4-DB	10.908	10.812	11.012	688.940	712.500	-3.3
2,4-DCAA	7.761	7.665	7.865	760.410	750.000	1.4
DICAMBA	7.963	7.867	8.067	732.640	705.000	3.9
DICHLORPROP	8.680	8.584	8.784	712.320	705.000	1.0
Dinoseb	11.287	11.191	11.391	668.880	705.000	-5.1