

### CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: P4602 SAS No.: P4602 SDG NO.: P4602

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL06 Date Analyzed: 11/11/2024

Lab Sample No.: HSTDCCC750 Data File : PS028406.D Time Analyzed: 16:33

| COMPOUND         | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|------------------|--------|-----------|--------|--------------------|-------------------|------|
|                  |        | FROM      | TO     |                    |                   |      |
| 2,4,5-T          | 9.547  | 9.452     | 9.652  | 727.150            | 712.500           | 2.1  |
| 2,4,5-TP(Silvex) | 9.255  | 9.160     | 9.360  | 727.880            | 712.500           | 2.2  |
| 2,4-D            | 8.375  | 8.279     | 8.479  | 706.140            | 705.000           | 0.2  |
| 2,4-DB           | 10.124 | 10.029    | 10.229 | 698.650            | 712.500           | -1.9 |
| 2,4-DCAA         | 7.246  | 7.150     | 7.350  | 749.600            | 750.000           | -0.1 |
| DICAMBA          | 7.434  | 7.338     | 7.538  | 723.500            | 705.000           | 2.6  |
| DICHLORPROP      | 8.144  | 8.048     | 8.248  | 710.450            | 705.000           | 0.8  |
| Dinoseb          | 11.339 | 11.245    | 11.445 | 715.780            | 705.000           | 1.5  |

### CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: P4602 SAS No.: P4602 SDG NO.: P4602

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL06 Date Analyzed: 11/11/2024

Lab Sample No.: HSTDCCC750 Data File : PS028406.D Time Analyzed: 16:33

| COMPOUND         | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|------------------|--------|-----------|--------|--------------------|-------------------|------|
|                  |        | FROM      | TO     |                    |                   |      |
| 2,4,5-T          | 10.328 | 10.233    | 10.433 | 676.610            | 712.500           | -5.0 |
| 2,4,5-TP(Silvex) | 9.908  | 9.811     | 10.011 | 676.600            | 712.500           | -5.0 |
| 2,4-D            | 9.000  | 8.905     | 9.105  | 665.620            | 705.000           | -5.6 |
| 2,4-DB           | 10.895 | 10.800    | 11.000 | 657.840            | 712.500           | -7.7 |
| 2,4-DCAA         | 7.752  | 7.655     | 7.855  | 697.880            | 750.000           | -6.9 |
| DICAMBA          | 7.953  | 7.856     | 8.056  | 674.960            | 705.000           | -4.3 |
| DICHLORPROP      | 8.669  | 8.573     | 8.773  | 656.080            | 705.000           | -6.9 |
| Dinoseb          | 11.276 | 11.180    | 11.380 | 665.940            | 705.000           | -5.5 |