

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: P4602 SAS No.: P4602 SDG NO.: P4602

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL07 Date Analyzed: 11/11/2024

Lab Sample No.: HSTDCCC750 Data File : PS028418.D Time Analyzed: 21:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.547	9.452	9.652	736.420	712.500	3.4
2,4,5-TP(Silvex)	9.254	9.160	9.360	736.480	712.500	3.4
2,4-D	8.374	8.279	8.479	714.440	705.000	1.3
2,4-DB	10.123	10.029	10.229	702.180	712.500	-1.4
2,4-DCAA	7.245	7.150	7.350	742.700	750.000	-1.0
DICAMBA	7.434	7.338	7.538	731.350	705.000	3.7
DICHLORPROP	8.143	8.048	8.248	717.160	705.000	1.7
Dinoseb	11.337	11.245	11.445	726.290	705.000	3.0

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: P4602 SAS No.: P4602 SDG NO.: P4602

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL07 Date Analyzed: 11/11/2024

Lab Sample No.: HSTDCCC750 Data File : PS028418.D Time Analyzed: 21:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.327	10.233	10.433	678.220	712.500	-4.8
2,4,5-TP(Silvex)	9.906	9.811	10.011	677.390	712.500	-4.9
2,4-D	9.000	8.905	9.105	636.580	705.000	-9.7
2,4-DB	10.894	10.800	11.000	666.670	712.500	-6.4
2,4-DCAA	7.751	7.655	7.855	698.440	750.000	-6.9
DICAMBA	7.952	7.856	8.056	673.550	705.000	-4.5
DICHLORPROP	8.669	8.573	8.773	656.760	705.000	-6.8
Dinoseb	11.274	11.180	11.380	666.370	705.000	-5.5