

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: P4602 SAS No.: P4602 SDG NO.: P4602

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL08 Date Analyzed: 11/12/2024

Lab Sample No.: HSTDCCC750 Data File : PS028430.D Time Analyzed: 02:11

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.545	9.452	9.652	724.840	712.500	1.7
2,4,5-TP(Silvex)	9.252	9.160	9.360	724.670	712.500	1.7
2,4-D	8.373	8.279	8.479	704.430	705.000	-0.1
2,4-DB	10.121	10.029	10.229	680.020	712.500	-4.6
2,4-DCAA	7.244	7.150	7.350	745.790	750.000	-0.6
DICAMBA	7.432	7.338	7.538	720.640	705.000	2.2
DICHLORPROP	8.141	8.048	8.248	709.240	705.000	0.6
Dinoseb	11.335	11.245	11.445	713.990	705.000	1.3

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: P4602 SAS No.: P4602 SDG NO.: P4602

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL08 Date Analyzed: 11/12/2024

Lab Sample No.: HSTDCCC750 Data File : PS028430.D Time Analyzed: 02:11

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.326	10.233	10.433	666.200	712.500	-6.5
2,4,5-TP(Silvex)	9.906	9.811	10.011	665.210	712.500	-6.6
2,4-D	8.999	8.905	9.105	614.680	705.000	-12.8
2,4-DB	10.894	10.800	11.000	644.510	712.500	-9.5
2,4-DCAA	7.750	7.655	7.855	684.690	750.000	-8.7
DICAMBA	7.951	7.856	8.056	662.040	705.000	-6.1
DICHLORPROP	8.668	8.573	8.773	643.140	705.000	-8.8
Dinoseb	11.274	11.180	11.380	655.310	705.000	-7.0