

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: P4602 SAS No.: P4602 SDG NO.: P4602

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL09 Date Analyzed: 11/12/2024

Lab Sample No.: HSTDCCC750 Data File : PS028442.D Time Analyzed: 06:59

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.542	9.452	9.652	726.580	712.500	2.0
2,4,5-TP(Silvex)	9.249	9.160	9.360	728.750	712.500	2.3
2,4-D	8.370	8.279	8.479	705.610	705.000	0.1
2,4-DB	10.118	10.029	10.229	661.620	712.500	-7.1
2,4-DCAA	7.242	7.150	7.350	746.620	750.000	-0.5
DICAMBA	7.430	7.338	7.538	722.500	705.000	2.5
DICHLORPROP	8.139	8.048	8.248	709.750	705.000	0.7
Dinoseb	11.332	11.245	11.445	716.220	705.000	1.6

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: P4602 SAS No.: P4602 SDG NO.: P4602

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL09 Date Analyzed: 11/12/2024

Lab Sample No.: HSTDCCC750 Data File : PS028442.D Time Analyzed: 06:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.323	10.233	10.433	659.950	712.500	-7.4
2,4,5-TP(Silvex)	9.902	9.811	10.011	654.880	712.500	-8.1
2,4-D	8.995	8.905	9.105	653.430	705.000	-7.3
2,4-DB	10.890	10.800	11.000	646.400	712.500	-9.3
2,4-DCAA	7.748	7.655	7.855	679.110	750.000	-9.5
DICAMBA	7.949	7.856	8.056	655.240	705.000	-7.1
DICHLORPROP	8.665	8.573	8.773	631.600	705.000	-10.4
Dinoseb	11.270	11.180	11.380	647.110	705.000	-8.2