

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: P4602 SAS No.: P4602 SDG NO.: P4602

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL10 Date Analyzed: 11/12/2024

Lab Sample No.: HSTDCCC750 Data File : PS028448.D Time Analyzed: 09:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.548	9.452	9.652	710.480	712.500	-0.3
2,4,5-TP(Silvex)	9.254	9.160	9.360	711.720	712.500	-0.1
2,4-D	8.374	8.279	8.479	689.320	705.000	-2.2
2,4-DB	10.124	10.029	10.229	688.180	712.500	-3.4
2,4-DCAA	7.244	7.150	7.350	824.620	750.000	9.9
DICAMBA	7.433	7.338	7.538	704.520	705.000	-0.1
DICHLORPROP	8.143	8.048	8.248	691.900	705.000	-1.9
Dinoseb	11.340	11.245	11.445	696.080	705.000	-1.3

CALIBRATION VERIFICATION SUMMARY

Contract: TETR16

Lab Code: CHEM Case No.: P4602 SAS No.: P4602 SDG NO.: P4602

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL10 Date Analyzed: 11/12/2024

Lab Sample No.: HSTDCCC750 Data File : PS028448.D Time Analyzed: 09:52

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.324	10.233	10.433	650.410	712.500	-8.7
2,4,5-TP(Silvex)	9.904	9.811	10.011	647.470	712.500	-9.1
2,4-D	8.997	8.905	9.105	631.840	705.000	-10.4
2,4-DB	10.892	10.800	11.000	622.480	712.500	-12.6
2,4-DCAA	7.748	7.655	7.855	662.720	750.000	-11.6
DICAMBA	7.949	7.856	8.056	640.850	705.000	-9.1
DICHLORPROP	8.666	8.573	8.773	618.400	705.000	-12.3
Dinoseb	11.273	11.180	11.380	633.710	705.000	-10.1