

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4602

Client: Tetra Tech, EMI

Analytical Method: 8151A

DataFile : PS028411.D

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	RPD
			Result	Result			Qual	RPD	Qual	Low	High	
Client Sample ID: P4602-41MS	CC0R6MS											
	DICAMBA	176	0	66.9	ug/Kg	38				10	112	
	DICHLORPROP	176	0	31.9	ug/Kg	18				10	113	
	2,4-D	176	0	73.5	ug/Kg	42				10	144	
	2,4,5-TP(Silvex)	176	0	23.0	ug/Kg	13				10	114	
	2,4,5-T	176	0	39.5	ug/Kg	22				10	115	
	2,4-DB	176	0	20.1	ug/Kg	11				10	140	
	Dinoseb	176	0	0	ug/Kg	0	*			10	118	