

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4602

Client: Tetra Tech, EMI

Analytical Method: 8151A

DataFile : PS028412.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: P4602-42MSD	CC0R6MSD											
	DICAMBA	176.1	0	67.9	ug/Kg	39		3		10	112	20
	DICHLORPROP	176.1	0	32.4	ug/Kg	18		0		10	113	20
	2,4-D	176.1	0	74.5	ug/Kg	42		0		10	144	20
	2,4,5-TP(Silvex)	176.1	0	23.3	ug/Kg	13		0		10	114	20
	2,4,5-T	176.1	0	40.1	ug/Kg	23		4		10	115	20
	2,4-DB	176.1	0	21.3	ug/Kg	12		9		10	140	20
	Dinoseb	176.1	0	0	ug/Kg	0	*	0		10	118	20