

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P4602

Client: Tetra Tech, EMI

Analytical Method: 8151A

DataFile : PS028437.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec Qual	RPD Qual	Low	Limits High	RPD
Client Sample ID: P4602-39MSD	CC0P0MSD									
	DICAMBA	179.1	0	42.0	ug/Kg	23	0	10	112	20
	DICHLORPROP	179.1	0	23.6	ug/Kg	13	0	10	113	20
	2,4-D	179.1	0	52.7	ug/Kg	29	0	10	144	20
	2,4,5-TP(Silvex)	179.1	0	11.4	ug/Kg	6	*	15	114	20
	2,4,5-T	179.1	0	21.3	ug/Kg	12	9	10	115	20
	2,4-DB	179.1	0	51.0	ug/Kg	28	7	10	140	20
	Dinoseb	179.1	0	13.9	ug/Kg	8	*	0	118	20