

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4602

Client: Tetra Tech, EMI

Analytical Method: 8151A **Datafile :** PS028231.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB164641BS	DICAMBA	166.5	181	ug/Kg	109				72	129		
	DICHLORPROP	166.5	185	ug/Kg	111				77	135		
	2,4-D	166.5	186	ug/Kg	112				65	144		
	2,4,5-TP(Silvex)	166.5	194	ug/Kg	117				74	146		
	2,4,5-T	166.5	193	ug/Kg	116				77	134		
	2,4-DB	166.5	188	ug/Kg	113				72	122		
	Dinoseb	166.5	197	ug/Kg	118				74	132		