



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4602

Client: Tetra Tech, EMI

Analytical Method: 8151A Datafile : PS028415.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB164725BS	DICAMBA	166.6	145	ug/Kg	87				72	129	
	DICHLORPROP	166.6	156	ug/Kg	94				77	135	
	2,4-D	166.6	170	ug/Kg	102				65	144	
	2,4,5-TP(Silvex)	166.6	167	ug/Kg	100				74	146	
	2,4,5-T	166.6	153	ug/Kg	92				77	134	
	2,4-DB	166.6	132	ug/Kg	79				72	122	
	Dinoseb	166.6	140	ug/Kg	84				74	132	