

SDG NARRATIVE

LAB NAME: Alliance Technical Group, LLC CASE: 51736 SDG: A4ED4 CONTRACT: 68HERH20D0011 LAB CODE: ACE LAB ORDER ID: P4604 MODIFICATION REF. NUMBER: NA

| Sample ID | EPA Sample ID | pН |
|-------------|----------------------|----|
| P4604-01 | A4ED4 | |
| P4604-02MS | A4ED4MS | |
| P4604-03MSD | A4ED4MSD | |

03 Soil samples were delivered to the laboratory intact on 10/29/2024.

Test requested on the Chain of Custody was Volatile Organic and Semivolatile Organic by Method SFAM01.1.

The temperature of the samples was measured using an I R Gun. The samples temperature was 3.0 degree Celsius for the samples received on 10/29/2024.

Low Volatiles:

The analysis performed on instrument MSVOA_W were done using GC column RXI-624SIL MS 30m 0.18mm 1.4 um. Cat#13868.

The analysis of VOC-SFAM was based on method SFAM01.1_LOW.

The Holding Times were met for all analysis. The Surrogate recoveries met the acceptable criteria. The Internal Standards Areas met the acceptable requirements. Instrument Performance Check met requirements. The Retention Times were met for all samples. The Tuning criteria met requirements. The MS {A4ED4MS} recovery met the requirements for all compounds. The MSD {A4ED4MSD} recovery met the requirements for all compounds. The RPD {A4ED4MSD} RPD met the requirements for all compounds. The Initial Calibration met the requirements.



2 of 4 The Continuing Calibration (VSTD025547) file ID VW030874.D met the requirements except for 1,2-Dichloroethane-d4 (25.9%). As per method, up to two target analyte in opening and closing CCV are allowed to exceed the %D values. Therefore no further corrective action was taken.

The blank analysis did not indicate the presence of lab contamination. The storage blank did not indicate the presence of lab contamination.

See **Manual Integration report f**or the manual integration information at the end of the case narrative.

Calculation:

Low/Med Level Soil/Sediment Calculation

Concentration in ug/Kg dry Weight basis) = (Ax)(Is)(Df)(Ais)(RRF)(Ws)(D)

Where,

Ax = Area for the compound to be measured Ais = Area for the specific internal standard Is = Amount of internal standard added in Nano grams (ng) RRF = Relative response factor of the calibration standard. Df = Dilution factor Ws= Weight of sample $D = \frac{100 - \%moisture}{100}$

No positive target compounds were detected in the sample.

Relative Response Factor = Dichlorodifluoromethane: RUN VW110424 for 2.5 ppb

RRF=Area of compound
Area of Internal StandardXConc. of Internal Standard
Conc. of Compound

 $RRF = \frac{13068}{434671} X \frac{25}{2.5}$

RRF= 0.301

Semivolatiles:

The samples were analyzed on instrument BNA_G using GC Column ZB-GR Semi Volatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGA.



Semis volatile Organic for soil sample was extracted by Method SFAM01.1 on 11/05/2024, The analysis of SVO-SFAM was based on method SFAM01.1_SVOC.

The Holding Times were met for all analysis. The Surrogate recoveries met the acceptable criteria. The Internal Standards Areas met the acceptable requirements. The Retention Times were acceptable for all samples. The MS {A4ED4MS} recovery met the requirements for all compounds. The MSD {A4ED4MSD} recovery met the requirements for all compounds. The RPD {A4ED4MSD} RPD met the requirements for all compounds The RPD {A4ED4MSD} RPD met the requirements for all compounds The Blank Spike for {PB164682BS} recoveries met the requirements for all compounds. The Blank analysis did not indicate the presence of lab contamination. The Tuning criteria met the requirements. The Initial Calibration met the requirements. The Continuous Calibration met the requirements.

Concentration of SOIL Sample:

Concentration ug/Kg,

(dry weight basis) = (Ax) (Is) (Vt) (DF) (GPC)

(Ais) (RRF) (Vi) (Wt) (D)

Where,

Ax = Area of the characteristic ion for the compound to be measured.

Ais = Area of the characteristic ion for the internal standard.

Is = Amount of internal standard injected in ng.

Vi = Volume of extract injected in microliters (uL)

Vt = Volume of concentrated extract in microliters (uL)

Wt = Weight of the original sample extracted in g

Df = Dilution factor

RRF = Mean Relative Response Factor determined from the initial calibration standard.

GPC = Vin = GPC factor (If no GPC is performed, GPC=1)

Vout = Volume of extract collected after GPC cleanup.

D= 100 - % moisture

100

No positive target compounds were detected in the samples.

RRF Calculation of standard 20 ppb for Naphthalene with G instrument for method 11/06/2024.



RRF= Area of compound / X Conc. of Internal Standard / Conc. of Compound /

= 267121/252685 X 20/20

= 1.057 (Reported RRF)

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature _____ Name: Nimisha Pandya.

Date: _____ Title: Document Control Officer.

4 of 4