ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092 NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEM	TECH PROJECT NUMBER: P4605	MATRIX: Solid		
METH	OD: 9045D,Lloyd Kahn			
1.	Blank Contamination - If yes, list compounds and concentra	NA ations in each blank:	NO ✓	YES
2.	Matrix Spike Duplicate Recoveries Met Criteria			✓
	If not met, list those compounds and their recoveries which range.	fall outside the acceptable		
	The Blank Spike met requirements for all samples.			
3.	Sample Duplicate Analysis Met QC Criteria			✓
	If not met, list those compounds and their recoveries which range.	fall outside the acceptable		
4.	Digestion Holding Time Met		✓	
	If not met, list number of days exceeded for each sample:			
	The Holding Times were met for all samples except for TAT2 of pH, for TAPHHA-SED02-102824-00-T2 of pH.for T00-T2 of pH.for TAPIAL2-SED02-102824-00-T2 of pH.for 102824-00-T2 of pH.for TAPLPR-SED11-102824-00-T2 or receive out of holding time.	APIAL2-SED01-102824- r TAPIAL3-SED01-		
ADDIT	IONAL COMMENTS:			
Lab has	used least representable sample weight for the sample P4605	-05 for TOC analysis. Therefore Lab	has repo	rted
he TOC	C result with "OR" qualifier.			
QA RE	VIEW	Date		