



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: P4605

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

RunNo.: LB133195

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.00	7	100	90-110	10/29/2024
Sample ID:	CCV1						
pH		pH	2.02	2.00	101	90-110	10/29/2024
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	10/29/2024

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: P4605

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

RunNo.: LB133199

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV1						
TOC		mg/L	1010	1000	101	90-110	08/07/2024
Sample ID:	CCV1						
TOC		mg/L	1000	1000	100	90-110	10/30/2024
Sample ID:	CCV2						
TOC		mg/L	1020	1000	102	90-110	10/30/2024
Sample ID:	CCV3						
TOC		mg/L	1060	1000	106	90-110	10/30/2024