

Matrix Spike Summary

Client:	Weston Solutions	SDG No.:	P4605
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Sample ID:	P4605-02
Client ID:	TAPIAL2-SED01-102824-00-T2MS	Percent Solids for Spike Sample:	85.8

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
TOC	mg/Kg	75-125	6270		5490		1000	1	78		10/30/2024

Matrix Spike Summary

Client:	Weston Solutions	SDG No.:	P4605
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Sample ID:	P4605-02
Client ID:	TAPIAL2-SED01-102824-00-T2MSD	Percent Solids for Spike Sample:	85.8

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
TOC	mg/Kg	75-125	6440		5490		1000	1	95		10/30/2024