

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P4606 SAS No.: P4606 SDG NO.: P4606

Continuing Calib Date: 11/04/2024 Initial Calibration Date(s): 10/15/2024 10/16/2024

Continuing Calib Time: 17:10 Initial Calibration Time(s): 18:27 02:36

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.52	5.52	5.42	5.62	0.00
Aroclor-1016-2	(2)	5.54	5.55	5.45	5.65	0.01
Aroclor-1016-3	(3)	5.61	5.61	5.51	5.71	0.00
Aroclor-1016-4	(4)	5.70	5.70	5.60	5.80	0.00
Aroclor-1016-5	(5)	6.00	6.00	5.90	6.10	0.00
Aroclor-1260-1	(1)	7.13	7.13	7.03	7.23	0.00
Aroclor-1260-2	(2)	7.38	7.38	7.28	7.48	0.00
Aroclor-1260-3	(3)	7.75	7.74	7.64	7.84	0.00
Aroclor-1260-4	(4)	7.97	7.97	7.87	8.07	0.00
Aroclor-1260-5	(5)	8.28	8.28	8.18	8.38	0.00
Tetrachloro-m-xylene		4.37	4.37	4.27	4.47	0.00
Decachlorobiphenyl		10.07	10.06	9.96	10.16	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P4606 SAS No.: P4606 SDG NO.: P4606

Continuing Calib Date: 11/04/2024 Initial Calibration Date(s): 10/15/2024 10/16/2024

Continuing Calib Time: 17:10 Initial Calibration Time(s): 18:27 02:36

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.72	4.73	4.63	4.83	0.01
Aroclor-1016-2	(2)	4.74	4.75	4.65	4.85	0.01
Aroclor-1016-3	(3)	4.92	4.92	4.82	5.02	0.00
Aroclor-1016-4	(4)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-5	(5)	5.17	5.18	5.08	5.28	0.01
Aroclor-1260-1	(1)	6.20	6.21	6.11	6.31	0.01
Aroclor-1260-2	(2)	6.39	6.39	6.29	6.49	0.00
Aroclor-1260-3	(3)	6.54	6.55	6.45	6.65	0.01
Aroclor-1260-4	(4)	7.02	7.02	6.92	7.12	0.01
Aroclor-1260-5	(5)	7.26	7.26	7.16	7.36	0.00
Tetrachloro-m-xylene		3.64	3.65	3.55	3.75	0.01
Decachlorobiphenyl		8.64	8.64	8.54	8.74	0.00