

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P4606 SAS No.: P4606 SDG NO.: P4606

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL01 Date Analyzed: 11/01/2024

Lab Sample No.: AR1660CCC500 Data File : PO107605.D Time Analyzed: 18:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.522	5.423	5.623	518.530	500.000	3.7
Aroclor-1016-2	5.544	5.445	5.645	505.780	500.000	1.2
Aroclor-1016-3	5.606	5.507	5.707	501.120	500.000	0.2
Aroclor-1016-4	5.703	5.604	5.804	526.240	500.000	5.2
Aroclor-1016-5	5.998	5.899	6.099	537.670	500.000	7.5
Aroclor-1260-1	7.128	7.026	7.226	505.110	500.000	1.0
Aroclor-1260-2	7.384	7.282	7.482	524.330	500.000	4.9
Aroclor-1260-3	7.747	7.643	7.843	531.800	500.000	6.4
Aroclor-1260-4	7.971	7.868	8.068	533.400	500.000	6.7
Aroclor-1260-5	8.285	8.181	8.381	516.940	500.000	3.4
Decachlorobiphenyl	10.068	9.958	10.158	47.490	50.000	-5.0
Tetrachloro-m-xylene	4.371	4.274	4.474	52.280	50.000	4.6

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P4606 SAS No.: P4606 SDG NO.: P4606

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL01 Date Analyzed: 11/01/2024

Lab Sample No.: AR1660CCC500 Data File : PO107605.D Time Analyzed: 18:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.725	4.626	4.826	513.740	500.000	2.7
Aroclor-1016-2	4.744	4.645	4.845	525.190	500.000	5.0
Aroclor-1016-3	4.920	4.821	5.021	508.170	500.000	1.6
Aroclor-1016-4	4.961	4.862	5.062	475.230	500.000	-5.0
Aroclor-1016-5	5.175	5.076	5.276	506.160	500.000	1.2
Aroclor-1260-1	6.206	6.107	6.307	489.530	500.000	-2.1
Aroclor-1260-2	6.394	6.294	6.494	519.700	500.000	3.9
Aroclor-1260-3	6.547	6.448	6.648	508.860	500.000	1.8
Aroclor-1260-4	7.018	6.918	7.118	503.610	500.000	0.7
Aroclor-1260-5	7.259	7.159	7.359	530.580	500.000	6.1
Decachlorobiphenyl	8.640	8.539	8.739	50.080	50.000	0.2
Tetrachloro-m-xylene	3.644	3.545	3.745	51.380	50.000	2.8