

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P4606 SAS No.: P4606 SDG NO.: P4606

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL02 Date Analyzed: 11/01/2024

Lab Sample No.: AR1660CCC500 Data File : PO107617.D Time Analyzed: 22:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.521	5.423	5.623	520.120	500.000	4.0
Aroclor-1016-2	5.544	5.445	5.645	502.650	500.000	0.5
Aroclor-1016-3	5.606	5.507	5.707	505.120	500.000	1.0
Aroclor-1016-4	5.703	5.604	5.804	534.940	500.000	7.0
Aroclor-1016-5	5.998	5.899	6.099	518.960	500.000	3.8
Aroclor-1260-1	7.127	7.026	7.226	502.860	500.000	0.6
Aroclor-1260-2	7.384	7.282	7.482	522.550	500.000	4.5
Aroclor-1260-3	7.746	7.643	7.843	528.660	500.000	5.7
Aroclor-1260-4	7.971	7.868	8.068	552.120	500.000	10.4
Aroclor-1260-5	8.285	8.181	8.381	522.950	500.000	4.6
Decachlorobiphenyl	10.071	9.958	10.158	47.880	50.000	-4.2
Tetrachloro-m-xylene	4.371	4.274	4.474	51.090	50.000	2.2

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P4606 SAS No.: P4606 SDG NO.: P4606

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL02 Date Analyzed: 11/01/2024

Lab Sample No.: AR1660CCC500 Data File : PO107617.D Time Analyzed: 22:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.723	4.626	4.826	507.760	500.000	1.6
Aroclor-1016-2	4.742	4.645	4.845	522.490	500.000	4.5
Aroclor-1016-3	4.918	4.821	5.021	502.610	500.000	0.5
Aroclor-1016-4	4.960	4.862	5.062	466.890	500.000	-6.6
Aroclor-1016-5	5.173	5.076	5.276	501.120	500.000	0.2
Aroclor-1260-1	6.204	6.107	6.307	491.540	500.000	-1.7
Aroclor-1260-2	6.392	6.294	6.494	520.280	500.000	4.1
Aroclor-1260-3	6.545	6.448	6.648	503.430	500.000	0.7
Aroclor-1260-4	7.016	6.918	7.118	504.530	500.000	0.9
Aroclor-1260-5	7.258	7.159	7.359	527.090	500.000	5.4
Decachlorobiphenyl	8.639	8.539	8.739	50.530	50.000	1.1
Tetrachloro-m-xylene	3.643	3.545	3.745	51.220	50.000	2.4