

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P4606 SAS No.: P4606 SDG NO.: P4606

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL04 Date Analyzed: 11/04/2024

Lab Sample No.: AR1660CCC500 Data File : PO107629.D Time Analyzed: 11:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.520	5.423	5.623	495.740	500.000	-0.9
Aroclor-1016-2	5.543	5.445	5.645	483.630	500.000	-3.3
Aroclor-1016-3	5.605	5.507	5.707	474.920	500.000	-5.0
Aroclor-1016-4	5.702	5.604	5.804	499.130	500.000	-0.2
Aroclor-1016-5	5.997	5.899	6.099	508.210	500.000	1.6
Aroclor-1260-1	7.126	7.026	7.226	509.270	500.000	1.9
Aroclor-1260-2	7.383	7.282	7.482	519.120	500.000	3.8
Aroclor-1260-3	7.745	7.643	7.843	522.220	500.000	4.4
Aroclor-1260-4	7.970	7.868	8.068	528.660	500.000	5.7
Aroclor-1260-5	8.283	8.181	8.381	520.190	500.000	4.0
Decachlorobiphenyl	10.065	9.958	10.158	48.540	50.000	-2.9
Tetrachloro-m-xylene	4.370	4.274	4.474	49.640	50.000	-0.7

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P4606 SAS No.: P4606 SDG NO.: P4606

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL04 Date Analyzed: 11/04/2024

Lab Sample No.: AR1660CCC500 Data File : PO107629.D Time Analyzed: 11:26

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.723	4.626	4.826	510.300	500.000	2.1
Aroclor-1016-2	4.742	4.645	4.845	522.520	500.000	4.5
Aroclor-1016-3	4.918	4.821	5.021	503.120	500.000	0.6
Aroclor-1016-4	4.959	4.862	5.062	468.200	500.000	-6.4
Aroclor-1016-5	5.172	5.076	5.276	500.090	500.000	0.0
Aroclor-1260-1	6.203	6.107	6.307	491.800	500.000	-1.6
Aroclor-1260-2	6.391	6.294	6.494	519.880	500.000	4.0
Aroclor-1260-3	6.545	6.448	6.648	498.840	500.000	-0.2
Aroclor-1260-4	7.015	6.918	7.118	504.220	500.000	0.8
Aroclor-1260-5	7.256	7.159	7.359	535.390	500.000	7.1
Decachlorobiphenyl	8.637	8.539	8.739	53.380	50.000	6.8
Tetrachloro-m-xylene	3.642	3.545	3.745	51.330	50.000	2.7