

Surrogate Summary

SDG No.: **P4606**

Client: **ATC Group Services LLC**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO107183.D	PIBLK-PO107183.D	Tetrachloro-m-xylene	1	20	22.5	112		60	140
		Decachlorobiphenyl	1	20	23.4	117		60	140
		Tetrachloro-m-xylene	2	20	22.1	110		60	140
		Decachlorobiphenyl	2	20	22.8	114		60	140
I.BLK-PO107610.D	PIBLK-PO107610.D	Tetrachloro-m-xylene	1	20	18.4	92		60	140
		Decachlorobiphenyl	1	20	18.5	92		60	140
		Tetrachloro-m-xylene	2	20	17.9	89		60	140
		Decachlorobiphenyl	2	20	18.3	92		60	140
PB164591BL	PB164591BL	Tetrachloro-m-xylene	1	20	22.0	110		32	144
		Decachlorobiphenyl	1	20	23.0	115		32	175
		Tetrachloro-m-xylene	2	20	21.7	108		32	144
		Decachlorobiphenyl	2	20	23.1	115		32	175
PB164591BS	PB164591BS	Tetrachloro-m-xylene	1	20	22.2	111		32	144
		Decachlorobiphenyl	1	20	22.8	114		32	175
		Tetrachloro-m-xylene	2	20	20.4	102		32	144
		Decachlorobiphenyl	2	20	23.1	115		32	175
I.BLK-PO107622.D	PIBLK-PO107622.D	Tetrachloro-m-xylene	1	20	18.4	92		60	140
		Decachlorobiphenyl	1	20	18.7	93		60	140
		Tetrachloro-m-xylene	2	20	17.9	89		60	140
		Decachlorobiphenyl	2	20	18.6	93		60	140
I.BLK-PO107625.D	PIBLK-PO107625.D	Tetrachloro-m-xylene	1	20	19.1	96		60	140
		Decachlorobiphenyl	1	20	21.2	106		60	140
		Tetrachloro-m-xylene	2	20	19.8	99		60	140
		Decachlorobiphenyl	2	20	22.4	112		60	140
P4606-01	1	Tetrachloro-m-xylene	1	20	0	0	*	32	144
		Decachlorobiphenyl	1	20	0	0	*	32	175
		Tetrachloro-m-xylene	2	20	0	0	*	32	144
		Decachlorobiphenyl	2	20	0	0	*	32	175
P4606-01DL	1DL	Tetrachloro-m-xylene	1	20	0	0	*	32	144
		Decachlorobiphenyl	1	20	0	0	*	32	175
		Tetrachloro-m-xylene	2	20	0	0	*	32	144
		Decachlorobiphenyl	2	20	0	0	*	32	175
I.BLK-PO107633.D	PIBLK-PO107633.D	Tetrachloro-m-xylene	1	20	19.4	97		60	140
		Decachlorobiphenyl	1	20	21.3	106		60	140
		Tetrachloro-m-xylene	2	20	19.4	97		60	140
		Decachlorobiphenyl	2	20	22.3	111		60	140