

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51716 MA No. : SDG No.: MC0PI2
Initial Calibration Verification Source : MP83271
Continuing Calibration Verification Source : MP83273
Run Batch: LB133469 Analytical Method: CVAA
Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV087				ID: CCV092				ID: CCV093		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury	4.0	3.6	90		5.0	5.1	103		5.0	101	

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51716 MA No. : SDG No.: MC0PI2
Initial Calibration Verification Source : MP83271
Continuing Calibration Verification Source : MP83273
Run Batch: LB133469 Analytical Method: CVAA
Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV094				ID: CCV095		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury					5.0	5.1	102		5.2	103	