


LB133469

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.3118e-004 slope

C= -9.6531e-002 intercept

Rho= 0.9998416

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.00	0.000	-0.035	-0.035	264	0.000	264					4.5
0.05	0.050					442	0				=
0.20	0.200	0.172	-0.028	1162	0.0 %	1162					14
2.50	2.500	2.540	0.040	11404	0.0 %	11404					2
5.00	5.000	5.117	0.117	22553	0.0 %	22553					2
7.50	7.500	7.408	-0.092	32461	0.0 %	32461					1
10.0	10.000	9.999	-0.001	43668	0.0 %	43668					0

LB133469
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	264	-		0 SFAM01.1	PPB	11/15/2024 14:10	Std
	0.2 S01	1162	-		0.2 SFAM01.1	PPB	11/15/2024 14:13	Std
	2.5 S02	11404	-		2.5 SFAM01.1	PPB	11/15/2024 14:15	Std
	5 S03	22553	-		5 SFAM01.1	PPB	11/15/2024 14:17	Std
	7.5 S04	32461	-		7.5 SFAM01.1	PPB	11/15/2024 14:19	Std
	10 S05	43668	-		10 SFAM01.1	PPB	11/15/2024 14:24	Std
ICV087	ICV087	16052	3.6143 -		SFAM01.1	PPB	11/15/2024 14:27	SMPL
ICB087	ICB087	148	-0.0623 -		SFAM01.1	PPB	11/15/2024 14:30	SMPL
CCV092	CCV092	22610	5.1304 -		SFAM01.1	PPB	11/15/2024 14:32	SMPL
CCB092	CCB092	136	-0.0651 -		SFAM01.1	PPB	11/15/2024 14:34	SMPL
PB165014BL	PBS014	179	-0.0551 -		SFAM01.1	PPB	11/15/2024 14:36	SMPL
P4584-03	MCC0P10	3271	0.6597 -		SFAM01.1	PPB	11/15/2024 14:39	SMPL
P4584-04	MCC0Q4	10445	2.3181 -		SFAM01.1	PPB	11/15/2024 14:41	SMPL
P4584-05	MCC0Q5	12327	2.7532 -		SFAM01.1	PPB	11/15/2024 14:43	SMPL
P4584-06	MCC0Q6	105	-0.0723 -		SFAM01.1	PPB	11/15/2024 14:46	SMPL
P4584-07	MCC0Q7	63540	14.5926 -		SFAM01.1	PPB	11/15/2024 14:48	SMPL
P4584-08	MCC0Q8	56966	13.0728 -		SFAM01.1	PPB	11/15/2024 14:50	SMPL
P4584-09	MCC0Q9	374	-0.0101 -		SFAM01.1	PPB	11/15/2024 14:52	SMPL
P4584-10	MCC0R0	2521	0.4863 -		SFAM01.1	PPB	11/15/2024 14:55	SMPL
P4584-11	MCC0R1	996	0.1337 -		SFAM01.1	PPB	11/15/2024 14:57	SMPL
P4584-12	MCC0R2	10520	2.3355 -		SFAM01.1	PPB	11/15/2024 14:59	SMPL
P4584-13	MCC0R3	18719	4.2309 -		SFAM01.1	PPB	11/15/2024 15:01	SMPL
P4584-14	MCC0R4	56520	12.9697 -		SFAM01.1	PPB	11/15/2024 15:04	SMPL
P4584-15	MCC0R5	-185	-0.1393 -		SFAM01.1	PPB	11/15/2024 15:06	SMPL
P4584-16	MCC0R6	1523	0.2556 -		SFAM01.1	PPB	11/15/2024 15:13	SMPL
P4584-17	MCC0R6D	1490	0.2479 -		SFAM01.1	PPB	11/15/2024 15:15	SMPL
P4584-18	MCC0R6S	12550	2.8048 -		SFAM01.1	PPB	11/15/2024 15:18	SMPL
P4584-19	MCC0R8	328	-0.0207 -		SFAM01.1	PPB	11/15/2024 15:20	SMPL
P4608-01	MCC0P0	226	-0.0443 -		SFAM01.1	PPB	11/15/2024 15:22	SMPL
CCV093	CCV093	22239	5.0446 -		SFAM01.1	PPB	11/15/2024 15:27	SMPL
CCB093	CCB093	60	-0.0827 -		SFAM01.1	PPB	11/15/2024 15:29	SMPL
P4608-03	MCC0POS	12801	2.8628 -		SFAM01.1	PPB	11/15/2024 15:31	SMPL
PB165015BL	PBS015	339	-0.0182 -		SFAM01.1	PPB	11/15/2024 15:34	SMPL
P4607-01	MCC0P12	27828	6.3367 -		SFAM01.1	PPB	11/15/2024 15:36	SMPL
P4607-02	MCC0P16	645	0.0526 -		SFAM01.1	PPB	11/15/2024 15:38	SMPL
P4607-03	MCC0P17	1438	0.2359 -		SFAM01.1	PPB	11/15/2024 15:40	SMPL
P4607-04	MCC0P18	33869	7.7333 -		SFAM01.1	PPB	11/15/2024 15:43	SMPL
P4607-05	MCC0P19	354	-0.0147 -		SFAM01.1	PPB	11/15/2024 15:45	SMPL
P4607-06	MCC0P0	326	-0.0212 -		SFAM01.1	PPB	11/15/2024 15:47	SMPL
P4607-07	MCC0P1	377	-0.0094 -		SFAM01.1	PPB	11/15/2024 15:49	SMPL
P4607-08	MCC0P2	946	0.1222 -		SFAM01.1	PPB	11/15/2024 15:52	SMPL
P4607-09	MCC0P3	1240	0.1901 -		SFAM01.1	PPB	11/15/2024 15:54	SMPL
P4607-10	MCC0P4	1023	0.14 -		SFAM01.1	PPB	11/15/2024 15:56	SMPL
P4607-11	MCC0P5	797	0.0877 -		SFAM01.1	PPB	11/15/2024 15:58	SMPL
P4607-12	MCC0P6	1248	0.192 -		SFAM01.1	PPB	11/15/2024 16:01	SMPL
P4607-13	MCC0P7	1465	0.2421 -		SFAM01.1	PPB	11/15/2024 16:03	SMPL

LB133469

INSTRUMENT ID : CV1

P4607-14	MCC0P8	3312	0.6691 -	SFAM01.1 PPB	11/15/2024 16:05 SMPL
P4607-15	MCC0P9	2832	0.5582 -	SFAM01.1 PPB	11/15/2024 16:08 SMPL
P4607-16	MCC0Q0	1662	0.2877 -	SFAM01.1 PPB	11/15/2024 16:10 SMPL
P4607-17	MCC0Q1	730	0.0722 -	SFAM01.1 PPB	11/15/2024 16:12 SMPL
P4607-18	MCC0Q2	732	0.0727 -	SFAM01.1 PPB	11/15/2024 16:14 SMPL
CCV094	CCV094	22451	5.0937 -	SFAM01.1 PPB	11/15/2024 16:17 SMPL
CCB094	CCB094	107	-0.0718 -	SFAM01.1 PPB	11/15/2024 16:19 SMPL
P4607-19	MCC0Q3	9937	2.2007 -	SFAM01.1 PPB	11/15/2024 16:21 SMPL
P4607-20	MCC0R7	645	0.0526 -	SFAM01.1 PPB	11/15/2024 16:23 SMPL
P4607-21	MCC0R7D	806	0.0898 -	SFAM01.1 PPB	11/15/2024 16:26 SMPL
P4607-22	MCC0R7S	12351	2.7588 -	SFAM01.1 PPB	11/15/2024 16:28 SMPL
P4608-02	MCC0P0D	421	0.0008 -	SFAM01.1 PPB	11/15/2024 16:30 SMPL
P4584-07DLX5	MCC0Q7	13316	2.9818 -	SFAM01.1 PPB	11/15/2024 16:32 SMPL
P4584-08DLX5	MCC0Q8	12168	2.7165 -	SFAM01.1 PPB	11/15/2024 16:35 SMPL
P4584-14DLX5	MCC0R4	11954	2.667 -	SFAM01.1 PPB	11/15/2024 16:37 SMPL
CCV095	CCV095	22783	5.1704 -	SFAM01.1 PPB	11/15/2024 16:39 SMPL
CCB095	CCB095	71	-0.0801 -	SFAM01.1 PPB	11/15/2024 16:42 SMPL