



Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	969	1000	97	90 - 110	P	11/01/2024	15:31	LB133257



Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	10.8	12.0	90	80 - 120	P	11/01/2024	15:35	LB133257

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4615
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4615 **SAS No.:** P4615
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	4830	5000	97	90 - 110	P	11/01/2024	16:03	LB133257
CCV02	Lead	5070	5000	102	90 - 110	P	11/01/2024	16:56	LB133257
CCV03	Lead	5090	5000	102	90 - 110	P	11/01/2024	17:57	LB133257
CCV04	Lead	4870	5000	97	90 - 110	P	11/01/2024	18:55	LB133257
CCV05	Lead	5340	5000	107	90 - 110	P	11/01/2024	19:18	LB133257



Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	979	1000	98	90 - 110	P	11/04/2024	17:28	LB133296

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4615
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4615 **SAS No.:** P4615
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	11.8	12.0	98	80 - 120	P	11/04/2024	17:37	LB133296

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4615
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4615 **SAS No.:** P4615
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	5120	5000	102	90 - 110	P	11/04/2024	18:11	LB133296
CCV02	Lead	5320	5000	106	90 - 110	P	11/04/2024	19:02	LB133296
CCV03	Lead	5320	5000	106	90 - 110	P	11/04/2024	19:46	LB133296
CCV04	Lead	5020	5000	100	90 - 110	P	11/04/2024	21:07	LB133296
CCV05	Lead	5020	5000	100	90 - 110	P	11/04/2024	21:47	LB133296

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4615
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4615 **SAS No.:** P4615
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	1020	1000	102	90 - 110	P	11/07/2024	13:49	LB133344



Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	11.4	12.0	95	80 - 120	P	11/07/2024	14:23	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4615

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4615 **SAS No.:** P4615

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	4920	5000	98	90 - 110	P	11/07/2024	15:01	LB133344
CCV02	Lead	4800	5000	96	90 - 110	P	11/07/2024	15:53	LB133344
CCV03	Lead	4970	5000	99	90 - 110	P	11/07/2024	16:45	LB133344
CCV04	Lead	5010	5000	100	90 - 110	P	11/07/2024	17:56	LB133344
CCV05	Lead	4880	5000	98	90 - 110	P	11/07/2024	18:32	LB133344
CCV06	Lead	4930	5000	99	90 - 110	P	11/07/2024	19:08	LB133344
CCV07	Lead	4850	5000	97	90 - 110	P	11/07/2024	19:58	LB133344
CCV08	Lead	4840	5000	97	90 - 110	P	11/07/2024	20:58	LB133344
CCV09	Lead	4780	5000	96	90 - 110	P	11/07/2024	21:51	LB133344
CCV10	Lead	4710	5000	94	90 - 110	P	11/07/2024	22:52	LB133344
CCV11	Lead	4800	5000	96	90 - 110	P	11/07/2024	23:43	LB133344
CCV12	Lead	4820	5000	96	90 - 110	P	11/08/2024	00:39	LB133344
CCV13	Lead	4870	5000	97	90 - 110	P	11/08/2024	00:57	LB133344