



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Portal Partners Tri-Venture

SDG No.: P4615

Contract: PORT06

Lab Code: CHEM

Case No.: P4615

SAS No.: P4615

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Lead	10.6	12.0	88	40 - 160	P	11/01/2024	15:44	LB133257
CRI01	Lead	11.7	12.0	98	40 - 160	P	11/04/2024	17:52	LB133296
CRI01	Lead	11.3	12.0	94	40 - 160	P	11/07/2024	14:42	LB133344