

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Portal Partners Tri-Venture **SDG No.:** P4615
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4615 **SAS No.:** P4615
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Lead	7.70			-12	12	11/01/2024	15:48	LB133257
ICSAB01	Lead	58.4	49.0	119	37	61	11/01/2024	15:58	LB133257
ICSA01	Lead	1.99			-12	12	11/04/2024	18:03	LB133296
ICSAB01	Lead	50.7	49.0	104	37	61	11/04/2024	18:07	LB133296
ICSA01	Lead	8.91			-12	12	11/07/2024	14:47	LB133344
ICSAB01	Lead	56.6	49.0	116	37	61	11/07/2024	14:51	LB133344