

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06  
 Lab Code: CHEM Case No.: P4615 SAS No.: P4615 SDG NO.: P4615  
 Lab File ID: VX043610.D Date Analyzed: 10/30/2024  
 Instrument ID: MSVOA\_X Time Analyzed: 09:26  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|----------------|---------------|-------|---------------|-------|---------------|--------|
| 12 HOUR STD    | 150073        | 5.54  | 269353        | 6.76  | 239151        | 10.06  |
| UPPER LIMIT    | 300146        | 6.044 | 538706        | 7.257 | 478302        | 10.555 |
| LOWER LIMIT    | 75036.5       | 5.044 | 134677        | 6.257 | 119576        | 9.555  |
| EPA SAMPLE NO. |               |       |               |       |               |        |
| TB-10292024    | 141860        | 5.55  | 257724        | 6.76  | 226949        | 10.06  |
| VX1030WBL01    | 124305        | 5.54  | 229429        | 6.76  | 204357        | 10.06  |
| VX1030WBS01    | 148306        | 5.54  | 264784        | 6.76  | 239021        | 10.06  |
| VX1030WBSD01   | 157679        | 5.55  | 281415        | 6.76  | 248802        | 10.06  |

IS1 = Pentafluorobenzene  
 IS2 = 1,4-Difluorobenzene  
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.