

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: PORT06
Lab Code: CHEM Case No.: P4615 SAS No.: P4615 SDG NO.: P4615
Lab File ID: VX043609.D BFB Injection Date: 10/30/2024
Instrument ID: MSVOA_X BFB Injection Time: 08:59
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	16.4
75	30.0 - 60.0% of mass 95	46.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	0.6 (0.8) 1
174	50.0 - 100.0% of mass 95	72
175	5.0 - 9.0% of mass 174	5.4 (7.5) 1
176	95.0 - 101.0% of mass 174	70.9 (98.5) 1
177	5.0 - 9.0% of mass 176	4.2 (6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX043610.D	10/30/2024	09:26
VX1030WBL01	VX1030WBL01	VX043612.D	10/30/2024	10:22
VX1030WBS01	VX1030WBS01	VX043613.D	10/30/2024	10:56
VX1030WBSD01	VX1030WBSD01	VX043614.D	10/30/2024	11:19
TB-10292024	P4615-04	VX043620.D	10/30/2024	13:38