

Continuing Calibration Report for SequenceID : FC110424AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067625.D

Aliphatic C9-C12	6816019.000	60.000	3.171	6.449	113600.317	135469.641	16.143
Aliphatic C12-C16	4594433.000	40.000	6.450	9.838	114860.825	131681.730	12.774
Aliphatic C16-C21	6776708.000	60.000	9.839	13.196	112945.133	127235.092	11.231
Aliphatic C21-C28	9587454.000	80.000	13.197	16.847	119843.175	120090.398	0.206
Aliphatic C28-C40	13067824.000	120.000	16.848	21.685	108898.533	101093.195	-7.721
Aliphatic EPH	40842438.000	360.000	3.171	21.685	113451.217	118799.912	4.502