

Continuing Calibration Report for SequenceID : FC110424AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067630.D

Aliphatic C9-C12	6934065.000	60.000	3.171	6.449	115567.750	135469.641	14.691
Aliphatic C12-C16	4673475.000	40.000	6.450	9.838	116836.875	131681.730	11.273
Aliphatic C16-C21	6853633.000	60.000	9.839	13.196	114227.217	127235.092	10.223
Aliphatic C21-C28	9678302.000	80.000	13.197	16.847	120978.775	120090.398	-0.740
Aliphatic C28-C40	13235928.000	120.000	16.848	21.685	110299.400	101093.195	-9.107
Aliphatic EPH	41375403.000	360.000	3.171	21.685	114931.675	118799.912	3.256