

Continuing Calibration Report for SequenceID : FD110424AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048640.D

Aromatic C10-C12	7057190.000	40.000	4.093	5.811	176429.750	176202.694	-0.129
Aromatic C12-C16	10560149.000	60.000	5.812	8.416	176002.483	177273.493	0.717
Aromatic C16-C21	12407458.000	80.000	8.417	12.676	155093.225	163832.550	5.334
Aromatic C21-C36	27916149.000	180.000	12.677	18.080	155089.717	152713.153	-1.556
Aromatic EPH	57940946.000	360.000	4.093	18.080	160947.072	161887.469	0.581