

Continuing Calibration Report for SequenceID : FD110424AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048657.D

Aromatic C10-C12	7232535.000	40.000	4.093	5.811	180813.375	176202.694	-2.617
Aromatic C12-C16	10879276.000	60.000	5.812	8.416	181321.267	177273.493	-2.283
Aromatic C16-C21	12923258.000	80.000	8.417	12.676	161540.725	163832.550	1.399
Aromatic C21-C36	29367307.000	180.000	12.677	18.080	163151.706	152713.153	-6.835
Aromatic EPH	60402376.000	360.000	4.093	18.080	167784.378	161887.469	-3.643