

Continuing Calibration Report for SequenceID : FD110424AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048665.D

Aromatic C10-C12	7508987.000	40.000	4.093	5.811	187724.675	176202.694	-6.539
Aromatic C12-C16	11220129.000	60.000	5.812	8.416	187002.150	177273.493	-5.488
Aromatic C16-C21	13232211.000	80.000	8.417	12.676	165402.638	163832.550	-0.958
Aromatic C21-C36	29153546.000	180.000	12.677	18.080	161964.144	152713.153	-6.058
Aromatic EPH	61114873.000	360.000	4.093	18.080	169763.536	161887.469	-4.865