

Continuing Calibration Report for SequenceID : FF110424AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF014905.D

Aromatic C10-C12	6096219.000	40.000	4.508	6.372	152405.475	156086.454	2.358
Aromatic C12-C16	9133431.000	60.000	6.373	9.064	152223.850	159619.601	4.633
Aromatic C16-C21	11072149.000	80.000	9.065	13.375	138401.863	155386.216	10.930
Aromatic C21-C36	23375895.000	180.000	13.376	18.817	129866.083	148444.550	12.515
Aromatic EPH	49677694.000	360.000	4.508	18.817	137993.594	152698.751	9.630