

Continuing Calibration Report for SequenceID : FF110424AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF014924.D

Aromatic C10-C12	6384117.000	40.000	4.508	6.372	159602.925	156086.454	-2.253
Aromatic C12-C16	9499667.000	60.000	6.373	9.064	158327.783	159619.601	0.809
Aromatic C16-C21	11432143.000	80.000	9.065	13.375	142901.788	155386.216	8.034
Aromatic C21-C36	24558734.000	180.000	13.376	18.817	136437.411	148444.550	8.089
Aromatic EPH	51874661.000	360.000	4.508	18.817	144096.281	152698.751	5.634