

Continuing Calibration Report for SequenceID : FG110424AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG014635.D

Aliphatic C9-C12	8683419.000	60.000	3.228	6.849	144723.650	131341.150	-10.189
Aliphatic C12-C16	5988867.000	40.000	6.850	10.292	149721.675	135821.940	-10.234
Aliphatic C16-C21	8883139.000	60.000	10.293	13.665	148052.317	141250.653	-4.815
Aliphatic C21-C28	10936782.000	80.000	13.666	17.329	136709.775	140581.144	2.754
Aliphatic C28-C40	12524334.000	120.000	17.330	22.266	104369.450	105686.976	1.247
Aliphatic EPH	47016541.000	360.000	3.228	22.266	130601.503	126992.540	-2.842