

Continuing Calibration Report for SequenceID : FG110424AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG014654.D

Aliphatic C9-C12	8864399.000	60.000	3.228	6.849	147739.983	131341.150	-12.486
Aliphatic C12-C16	6060981.000	40.000	6.850	10.292	151524.525	135821.940	-11.561
Aliphatic C16-C21	8934466.000	60.000	10.293	13.665	148907.767	141250.653	-5.421
Aliphatic C21-C28	10963583.000	80.000	13.666	17.329	137044.788	140581.144	2.516
Aliphatic C28-C40	12678278.000	120.000	17.330	22.266	105652.317	105686.976	0.033
Aliphatic EPH	47501707.000	360.000	3.228	22.266	131949.186	126992.540	-3.903