

WATER EPH MATRIX SPIKE/MATRIX SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Weston Solutions, Inc.
Lab Code: CHEM **Cas No:** P4620 **SAS No :** P4620 **SDG No:** P4620
Sample No : P4620-26 **Datafile:** FF014917.D
Client ID : P001-CF16-01MS

COMPOUND	SPIKE ADDED ug/l	SAMPLE CONCENTRATION ug/l	MS/MSD CONCENTRATION ug/l	% REC	Qual	QC LIMITS
Aromatic C10-C12	200	0	155	78		(40-140)
Aromatic C12-C16	300	0	250	83		(40-140)
Aromatic C16-C21	500	43.7	444	80		(40-140)
Aromatic C21-C36	800	0	637	80		(40-140)