

WATER EPH MATRIX SPIKE/MATRIX SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Weston Solutions, Inc.
Lab Code: CHEM **Cas No:** P4620 **SAS No :** P4620 **SDG No:** P4620
Sample No : P4620-26 **Datafile:** FG014647.D
Client ID : P001-CF16-01MS

COMPOUND	SPIKE ADDED ug/l	SAMPLE CONCENTRATION ug/l	MS/MSD CONCENTRATION ug/l	% REC	Qual	QC LIMITS
Aliphatic C9-C12	300	0	243	81		(40-140)
Aliphatic C12-C16	200	0	191	96		(40-140)
Aliphatic C16-C21	300	0	282	94		(40-140)
Aliphatic C21-C28	400	0	347	87		(40-140)
Aliphatic C28-C40	600	0	577	96		(40-140)