

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Weston Solutions, Inc.
Lab Code: CHEM **Cas No:** P4620 **SAS No :** P4620 **SDG No:** P4620
Sample No : PB164637BS **Datafile:** FF014921.D
Client ID : PB164637BS

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	QC LIMITS
Aromatic C10-C12	200	113	56		(40-140)
Aromatic C12-C16	300	201	67		(40-140)
Aromatic C16-C21	500	364	73		(40-140)
Aromatic C21-C36	800	587	73		(40-140)