

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: P4620 SAS No.: P4620 SDG NO.: P4620

Continuing Calib Date: 10/31/2024 Initial Calibration Date(s): 10/08/2024 10/08/2024

Continuing Calib Time: 14:53 Initial Calibration Time(s): 16:30 23:46

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	6.78	6.77	6.67	6.87	-0.01
Aroclor-1254-2	(2)	7.00	6.99	6.89	7.09	-0.01
Aroclor-1254-3	(3)	7.36	7.35	7.25	7.45	-0.01
Aroclor-1254-4	(4)	7.65	7.64	7.54	7.74	-0.01
Aroclor-1254-5	(5)	8.06	8.06	7.96	8.16	0.00
Tetrachloro-m-xylene		4.76	4.75	4.65	4.85	-0.01
Decachlorobiphenyl		10.68	10.67	10.57	10.77	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: P4620 SAS No.: P4620 SDG NO.: P4620

Continuing Calib Date: 10/31/2024 Initial Calibration Date(s): 10/08/2024 10/08/2024

Continuing Calib Time: 14:53 Initial Calibration Time(s): 16:30 23:46

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.97	5.98	5.88	6.08	0.01
Aroclor-1254-2 (2)	6.12	6.12	6.02	6.22	0.00
Aroclor-1254-3 (3)	6.53	6.53	6.43	6.63	0.00
Aroclor-1254-4 (4)	6.76	6.76	6.66	6.86	0.00
Aroclor-1254-5 (5)	7.18	7.18	7.08	7.28	0.00
Tetrachloro-m-xylene	4.05	4.05	3.95	4.15	0.00
Decachlorobiphenyl	9.21	9.22	9.12	9.32	0.01